









## F&P 950 System

### Blender Transition Kits Specifications

PRODUCT	OJR410B 	OJR412B 	OJR414B 
NASAL CANNULA SIZE	<b>XS</b>	<b>S</b>	<b>M</b>
Description	F&P Optiflow™ Junior 2 Blender Transition Kit – XS	F&P Optiflow Junior 2 Blender Transition Kit – S	F&P Optiflow Junior 2 Blender Transition Kit – M
Weight (per kit)	31.2 g	31.6 g	33.8 g
<b>FLOW RATES* (L/min) - F&amp;P 950™</b>			
	0.5-10	0.5-10	0.5-11
<b>COMPATIBLE SPARES</b>			
F&P Wigglepads™ 2	 WJR110	 WJR112	
<b>PRODUCT SPECIFICATIONS</b>			
Compatible circuits	950N60, 950N61, 950N62, 950N80, 950N81		
Connection	Easy-click connector, adaptor		
Compatible humidifier	F&P 950		
Quantity	Box of 5		
Box components	F&P Optiflow Junior 2 nasal cannula, proprietary adaptor, OJR215 pressure relief manifold, user instructions		

\* Flow rates above describe technical capability of the product when used at sea level. Ensure clinical judgement is used when prescribing flow rates.

**F&P** 950 System Blender Transition Kits | **SPECIFICATIONS**

<b>PERFORMANCE SPECIFICATIONS</b>	
Ambient range	20–26 °C / 68–79 °F
Usage	Single patient use
Duration of use	7 days
Recommended gas source	Medical gas
Shelf life	3 years
Nominal relief pressure	40 cmH <sub>2</sub> O
<b>COMPONENTS AND COMPOSITION</b>	
Cannula predominant materials	Thermoplastic Elastomer; Hydrocolloid; ABS; Stainless steel
Adaptor predominant materials	Polypropylene
Manifold predominant materials	Thermoplastic Elastomer White; ABS; Stainless steel; Silicon rubber
Materials not present	Not manufactured with natural rubber latex, PVC, or Phthalates (DEHP, DBP, BBP)
Manufacturing mode	Produced in a controlled working environment
Disposal	Incineration or according to hospital protocol for cannula, manifold and adapter; Clamshell packaging and label recyclable PET
<b>REGULATORY</b>	
Classification	AU-IIa, EU-IIa, Canada-II, US-II
Country of origin	New Zealand
Notified body	TÜV SÜD Product Services GmbH CE0123
Standard	Medical tapers used comply with ISO 5356-1

Please note that the information in this specifications sheet (including product information and images) is summarized and provided for illustrative purposes only. Please refer to the relevant user instructions for more information and confirm details with your local Fisher & Paykel Healthcare representative prior to placing an order. Information subject to change without notice.

F&P 950, Optiflow and Wigglepads are trademarks of Fisher & Paykel Healthcare Limited.