






**F&P** 950 System

Ventilator Transition Kits Specifications



| PRODUCT                           | OJR410VT   | OJR412VT  | OJR414VT   |
|-----------------------------------|--|---|--|
| NASAL CANNULA SIZE                | XS   | S   | M  |
| Description                       | F&P Optiflow™ Junior 2 Ventilator Transition Kit – XS  | F&P Optiflow™ Junior 2 Ventilator Transition Kit – S  | F&P Optiflow™ Junior 2 Ventilator Transition Kit – M |
| Weight (per kit)                  | 8.9 g  | 9.3 g   | 11.5 g   |
| <b>COMPATIBLE SPARES</b>          |  |   |  |
| F&P Wigglepad™ 2                  |  WJR110                            |  WJR112 |  |
| <b>PRODUCT SPECIFICATIONS</b>     |  |   |  |
| Compatible circuits               | 950N80, 950N81   |   |  |
| Connection                        | Easy-click connector   |   |  |
| Compatible humidifier             | F&P 950 (in invasive mode)   |   |  |
| Quantity                          | Box of 5   |   |  |
| Box components                    | F&P Optiflow Junior nasal cannula, proprietary adaptor, user instructions  |   |  |
| <b>PERFORMANCE SPECIFICATIONS</b> |  |   |  |
| Ambient range °C                  | 20-26 °C / 68-79 °F  |   |  |
| Usage                             | Single patient use   |   |  |
| Duration of use                   | 7 days   |   |  |
| Recommended gas source            | Medical gas  |   |  |
| Shelf life                        | 3 years  |   |  |
| <b>COMPONENTS AND COMPOSITION</b> |  |   |  |
| Predominant materials             | Thermoplastic Elastomer; Hydrocolloid; ABS; Stainless steel   Adaptor: Polypropylene                                 |   |  |
| Materials not present             | Not manufactured with natural rubber latex, PVC or Phthalates (DEHP, DBP, BBP)                                       |   |  |
| Manufacturing mode                | Produced in a controlled working environment   |   |  |
| Disposal                          | Incineration or according to hospital protocol for cannula and adapter; Clamshell packaging and label recyclable PET |   |  |
| <b>REGULATORY</b>                 |  |   |  |
| Classification                    | AU-IIa, EU-IIa, Canada-II, US-I  |   |  |
| Country of origin                 | New Zealand  |   |  |
| Notified body                     | TÜV SÜD Product Services GmbH CE0123   |   |  |

## Common ventilators

| PRODUCT  | OJR410VT  | OJR412VT  | OJR414VT  |
|--|--|--|--|
| NASAL CANNULA SIZE   | XS   | S  | M  |
| <b>MAX FLOW RATES* (L/min)</b>   |  |  |  |
| <b>Draeger Babylog VN500 Ventilator</b><br><b>Draeger Evita Infinity V500 Ventilator</b><br>Pressure limit 30 cmH <sub>2</sub> O<br>Ventilator mode O <sub>2</sub> therapy | 10   | 11   | 12   |
| <b>Maquet Servo-U Ventilator</b><br>Pressure limit 50 cmH <sub>2</sub> O<br>Ventilator mode High flow  | 9  | 10   | 10   |
| <b>Hamilton-S1 Ventilator</b><br><b>Hamilton-G5 Ventilator</b><br>Pressure limit 60 cmH <sub>2</sub> O<br>Ventilator mode High flow O <sub>2</sub>                         | 11   | 12   | 13   |

\* Flow rates above describe technical capability of the product when used at sea level.  
 Ensure clinical judgement is used when prescribing flow rates.  
 Maximum flow rates achieved are ventilator dependent.  
 Refer to the manufacturer's instructions for correct usage.

Please note that the information in this specifications sheet (including product information and images) is summarized and provided for illustrative purposes only. Please refer to the relevant user instructions for more information and confirm details with your local Fisher & Paykel Healthcare representative prior to placing an order. Information subject to change without notice.