


Mask Specifications



F&P Visairo™ Hospital Under Nose Mask	Product Order Codes		
	Non-vented Hospital Under Nose Mask Anti-asphyxiation Valve Version	Non-vented Hospital Under Nose Mask Standard Elbow Version	Vented Hospital Under Nose Mask Anti-asphyxiation Valve Version
F&P Visairo - Size A F&P Visairo - Size B F&P Visairo - Size C	RT075A RT075B RT075C	RT076A RT076B RT076C	RT077A RT077B RT077C
Recommended F&P 850™ Circuit*	RT219/RT319	RT280/RT380/RT481	RT219/RT319
Recommended F&P 950™ Circuit*	950A60/950A61	950A80/950A81/950A82	950A60/950A61
Circuit setup	Single limb with exhalation port	Dual limb	Single limb
Resistance to flow through mask	0.23 cmH ₂ O @ 50 L/min, 0.51 cmH ₂ O @ 100 L/min	0.07 cmH ₂ O @ 50 L/min, 0.33 cmH ₂ O @ 100 L/min	0.26 cmH ₂ O @ 50 L/min, 0.63 cmH ₂ O @ 100 L/min
Interface connections	ISO 5356-1 Conical Connectors		
Operating pressures	4 - 40 cmH ₂ O		
Box quantity	10		
Intended use	Single patient use, adult patient (>30kg), maximum use 14 days		
Materials	Thermoplastic Elastomer, Silicone, Acetal, Polycarbonate, Polyurethane Foam, Nylon, Nylon Spandex DEHP free; BPA free; not made with natural rubber latex		
Biocompatibility	Meets standards: ISO 10993-1:2018, ISO 18562-1:2017		
Packaging Materials	Outer: Carboard box; Inner: Polyester/Polyethylene laminate; Sizing guide: Cardboard		
Manufacturing Mode	Non invasive device; medically clean		
Destruction	Incineration, or according to hospital protocol		
Classification	USA - Class II; EU - Class IIa; AU - Class IIa		

Always refer to full user instructions supplied with the product. *Products mentioned may not be available in your region.